SD Card, real time data logger, + type K/J Temp., Patent

HUMIDITY/TEMP. METER

Model : HT-3007SD

www.yalab.com.tw

%RH POWER REC ESC ENTER LOGGER SET Sampling Time check check SD Card Data Logger Humidity / Temp. Meter DEW point, WET bulb Type K/J Thermometer LIFLutron HT-3007SD





ISO-9001, CE, IEC1010

02-2389-0101





The Art of Measurement

SD Card real time data logger HUMIDITY/TEMP. METER, + type K/J Temp. Model: HT-3007SD

FEATUR

| FLATOR |
|--|
| *ESumidity/Temp., Dew point (Humidity), Wet bulb |
| (Humidity), Type K/J thermometer, 4 kinds |
| measurement are combined into one meter, intelligent |
| and professional. |
| * Type K, Type J thermocouple thermometer. |
| * Fast humidity measuring response time. |
| * Real time SD memory card Datalogger, it Built-in Clock |
| and Calendar, real time data recorder, sampling time set |
| from 1 second to 3600 seconds. |
| * Manual datalogger is available (set the sampling |
| time to 0), during execute the manual datalogger |
| function, it can set the different position (location) No. |
| (position 1 to position 99). |
| * Innovation and easy operation, computer is not |
| need to setup extra software, after execute |
| datalogger, just take away the SD card from the |
| meter and plug in the SD card into the computer, |
| it can down load the all the measured value with |
| the time information (year/month/date/ |
| hour/minute/second) to the Excel directly, then |
| user can make the further data or graphic |
| analysis by themselves. |
| * SD card capacity : 1 GB to 16 GB. |
| * LCD with green light backlight, easy reading. |
| * Can default auto power off or manual power off. |
| * Data hold, record max. and min. reading. |
| * Microcomputer circuit, high accuracy. |
| * Power by UM3/AA (1.5 V) x 6 batteries or DC 9V adapter. |
| * RS232/USB PC COMPUTER interface. |
| * Separate probe, easy for operation of different |
| measurement environment. |
| |

| GENERAL SPEC | IFICATION | IS | |
|---------------|--|--------------------------------------|--|
| Circuit | Custom on | e-chip of microprocessor LSI | |
| | circuit. | | |
| Display | LCD size : 52 mm x 38 mm | | |
| | LCD with green backlight (ON/OFF). | | |
| Measurement | Humidity/Temp. : %RH/℃ or °F. | | |
| Unit | Dew point (Humidity) : ℃ or °F. | | |
| | Wet bulb (Humidity): $^{\circ}$ C or $^{\circ}$ F. | | |
| | Type K/J thermometer : $^{\circ}$ C or $^{\circ}$ F. | | |
| Datalogger | Auto 1 second to 3600 seconds | | |
| Sampling Time | | @ Sampling time can set to 1 second, | |
| Setting range | | but memory data may loss. | |
| | Manual | Push the data logger button | |
| | | once will save data onetime. | |
| | | @ Set the sampling time to | |
| | | 0 second. | |
| | | @ Manual mode, can also select the | |
| | | 1 to 99 position (Location) no. | |
| Memory Card | SD memor | y card. 1 G to 16 G. | |
| Advanced | * Set clock t | time (Year/Month/Date, | |
| setting | Hour/Min | ute/ Second) | |
| | * Decimal p | oint of SD card setting | |
| | * Auto powe | er OFF management | |
| | * Set beep Sound ON/OFF | | |
| | * Set thermometer type to Type K or Type J | | |
| | * Set temperature unit to °C or °F | | |
| | * Set sampling time | | |
| | * SD memory card Format | | |
| Temperature | Automatic temp. compensation for the | | |
| Compensation | humidity function and the type K/J | | |
| | thermometer. | | |
| Data Hold | Freeze the display reading. | | |
| Memory Recall | Maximum & Minimum value. | | |
| Sampling Time | Approx. 1 | second. | |
| of Display | | | |
| Data Output | RS 232/USB PC computer interface. | | |
| | * Connect the optional RS232 cable | | |
| | UPCB-02 will get the RS232 plug. | | |
| | | the optional USB cable | |
| | | vill get the USB plug. | |
| Operating | 0 to 50 °⊂. | | |
| Temperature | | | |
| Operating | Less than 85% R.H. | | |
| Humidity | | | |
| | | | |
| | <u> </u> | | |

www.YaLAB.com.tw 02-2389-0101

| Power Supply | * Alkaline or heavy duty DC 1.5 V battery | | |
|---------------|---|--|--|
| Fower Suppry | (UM3, AA) x 6 PCs, or equivalent. | | |
| | * DC 9V adapter input. (AC/DC power | | |
| | adapter is optional). | | |
| Power Current | Normal operation (w/o SD card save | | |
| | data and LCD Backlight is OFF) : | | |
| | Approx. DC 3.5 mA. | | |
| | When SD card save the data but and | | |
| | LCD Backlight is OFF) : Approx. DC 28 mA. | | |
| | | | |
| | * If LCD backlight on, the power | | |
| | consumption will increase approx. | | |
| | 12 mA. | | |
| Weight | 489 g/1.08 LB. | | |
| Dimension | 177 x 68 x 45 mm | | |
| | (7.0 x 2.7x 1.9 inch) | | |
| Accessories | * Instruction manual 1 PC | | |
| Included | * Humidity/Temp. probe1 PC | | |
| | * Soft carrying case (CA-05A) 1 PC | | |
| Optional | * SD Card (4 GB) | | |
| Accessories | * Type K thermocoupleprobe. | | |
| | TP-1, TP-02A, TP-03, TP-04, TP-05 | | |
| | * AC to DC 9V adapter. | | |
| | * USB cable, USB-01. | | |
| | * RS232 cable, UPCB-02. | | |
| | * Hard carrying case, CA-06 | | |
| | * Data Acquisition software, SW-U801-WIN. | | |
| | * Excel Data Acquisition software,SW-E802 | | |

ELECTRICAL SPECIFICATIONS (23± 5 °C)

Humidity/ Temperature

| | Range | 5 % to 95 % R.H. | |
|-------------|------------|--------------------------------|--|
| Humidity | Resolution | 0.1 % R.H. | |
| , | Accuracy | ≧70% RH : | |
| | | ± (3% reading + 1% RH). | |
| | | < 70% RH : | |
| ± 3% RH. | | ± 3% RH. | |
| | Range | 0 °C to 50 °C,32 °F to 122 °F. | |
| Temperature | Resolution | 0.1 degree | |
| | Accuracy | °C ± 0.8 °C. | |
| | - | °F ± 1.5 °F. | |

Dew Point (Humidity)

| °C | Range | -25.3 ℃ to 48.9 ℃ |
|----|------------|-----------------------|
| | Resolution | 0.1 °C |
| °F | Range | -13.5 °F to 120.1 °F. |
| | Resolution | 0.1 °F. |
| | | |

Remark :

* Dew Point display value is calculated from the

Humidity/Temp. measurement automatically.

* The Dew Point accuracy is sum accuracy value of Humidity & Temperature measurement..

Wet bulb (Humidity)

| °C | Range | -21.6 ℃ to 50.0 ℃ | |
|---|------------|----------------------|--|
| | Resolution | 0.1 °C | |
| °F | Range | -6.9 °F to 122.0 °F. | |
| | Resolution | 0.1 °F. | |
| Remark : | | | |
| * Wet bulb display value is calculated from the | | | |
| Humidity/Temp. measurement automatically. | | | |
| * The Dew Point accuracy is sum accuracy value of | | | |

Humidity & Temperature measurement...

Type K/J thermometer

| Type R/S diefinienee | | | |
|----------------------|------------|--------------------|----------------------|
| Sensor | Resolution | Range | Accuracy |
| Туре | | | |
| Туре К | 0.1 °C | -50.0 to 1300.0 ℃ | ±(0.4 % + 0.5 °C) |
| | | -50.1 to -100.0 °C | ± (0.4 % + 1 °C) |
| | 0.1 °F | -58.0 to 2372.0 °F | ± (0.4 % + 1 °F) |
| | | -58.1 to -148.0 °F | ± (0.4 % + 1.8 °F) |
| Type J | 0.1 °C | -50.0 to 1200.0 °C | ± (0.4 % + 0.5 °C) |
| | | -50.1 to -100.0 °C | ± (0.4 % + 1 °C) |
| | 0.1 °F | -58.0 to 2192.0 °F | ± (0.4 % + 1 °F) |
| | | -58.1 to -148.0 °F | ± (0.4 % + 1.8 °F) |

PATENT

CHINA: ZL 2008 2 0189918.5 ZL 2008 2 0189917.0 U.S.A. Pending

Germany : Nr. 20 2008 016 337.4 JAPAN : 3151214 * Appearance and specifications listed in this brochure are subject to change without notice. ľ