

SD card real time data recorder

**X, Y, Z, 3 Axis**

# VIBRATION METER

Model : BVB-8217SD

ISO-9001, CE, IEC1010

[www.yalab.com.tw](http://www.yalab.com.tw) 02-2389-0101



**Lutron**

[www.YaLAB.com.tw](http://www.YaLAB.com.tw)

*The Art of Measurement*

# SD Card real time data recorder X, Y, Z, 3 Axis **VIBRATION METER**

Model : BVB-8217SD

[www.YaLAB.com.tw](http://www.YaLAB.com.tw) 02-2389-0101

## FEATURE

* X, Y, Z, 3 Axis vibration meter.
* 3 channels vibration meter, use SD card to save the data along with time information, ( year, month, date, hour, minute, second ) into the SD memory card and can be downloaded to the Excel, extra software is no need.
* Show X, Y, Z ( CH1 to CH3 ) and 3 Axis vibration values in the same LCD.
* Applications for industrial vibration monitoring : All industrial machinery vibrates. The level of vibration is a useful guide to machine condition. Poor balance, misalignment & looseness of the structure will cause the vibration level increase, it is a sure sign that the maintenance is needed.
* Display no. : X, Y, X and 3 Axis.
* Frequency range 10 Hz - 1 kHz, sensitivity relative meet ISO 2954.
* Professional vibration meter supply with onevibration sensor ( X axis sensor ) and magnetic base, full set.
* Metric & Imperial display unit
* Acceleration, Velocity, Displacement measurement. Acceleration : 200 m/s <sup>2</sup> . Velocity : 200 mm/s. Displacement ( p-p ) : 2 mm.
* RMS, Max hold, Peak value measurement.
* Max. Hold reset button, Zero button.
* Wide frequency range.
* Data hold button to freeze the desired reading.
* Memory function to record maximum and minimum reading with recall.
* Separate vibration probe with magnetic base, easy operation.
* Real time SD memory card Datalogger, it Built-in Clock and Calendar, real time data recorder , sampling time set from 1 second to 3600 seconds.
* Manual datalogger is available ( set the sampling time to 0 ), during execute the manual datalogger function, it can set the different position ( location ) No. ( position 1 to position 99 ).
* Innovation and easy operation, computer is not need to setup extra software, after execute datalogger, just take away the SD card from the meter and plug in the SD card into the computer, it can down load the all the measured value with the time information ( year/month/date/ hour/minute/second ) to the Excel directly, then user can make the further data or graphic analysis by themselves.
* SD card capacity : 1 GB to 16 GB.
* LCD with green light backlight, easy reading.
* Can default auto power off or manual power off.
* Data hold, record max. and min. reading.
* Microcomputer circuit, high accuracy.
* Power by UM3/AA ( 1.5 V ) x 8 batteries or DC 9V adapter.
* RS232/USB PC COMPUTER interface.
* Include 1 PC sensor ( X sensor ) vibration set, VB-83. Extra vibration sensors ( X, Y sensor ) set, VB-83 can be ordered. When changing the VB-83, it is not necessary to make the new calibration again.

## GENERAL SPECIFICATION

Circuit	Custom one-chip of microprocessor LSI circuit.
Display	LCD size : 82 mm x 61 mm. * with green color backlight.

Channels	Show X, Y, Z ( CH1 to CH3 ) and 3 Axis vibration values in the same LCD.		
Measurement	Velocity, Acceleration, Displacement		
Function	Acceleration, Velocity :RMS, Peak, Max Hold. Displacement :p-p ( peak-peak ), Max Hold p-p.		
Unit	Measurement	Metric	Imperial
	Acceleration	meter/s <sup>2</sup> , g	ft/s <sup>2</sup> ,
	Velocity	mm/s, cm/s	inch/s
	Displacement	mm	inch
Frequency range	10 Hz to 1 KHz * Sensitivity relative during the the frequency range meet ISO 2954 Refer to table 1, page 30.		
Circuit	Exclusive microcomputer circuit.		
Peak Measurement	Acceleration, Velocity :To measure and update the peak value. Displacement :To measure and update the peak to peak ( p-p ) value.		
Max Hold Measurement	Acceleration, Velocity :To measure and update the max. peak value.		
	Displacement :To measure and update the max. peak to peak ( p-p ) value.		
Zero Button	Under Acceleration ( RMS ) measurement, sensor motionless , press two Buttons ( 3-5, 3-7, Fig. 1 ) >3 seconds.		
Max. Hold Reset Button	Under Max. hold measurement, press two Buttons ( 3-5, 3-7, Fig. 1 ) >3 seconds.		
Datalogger Sampling Time Setting range	Auto	1 second to 3600 seconds @ Sampling time can set to 1 second, but memory data may loss.	
	Manual	Push the data logger button once will save data one time. @ Set the sampling time to 0 second. @ Manual mode, can also select the1 to 99 position ( Location ) no.	
Data error no.	≤ 0.1 % no. of total saved data typically.		
Memory Card	SD memory card 1 GB to 16 GB.		
Advanced setting	<ul style="list-style-type: none"> <li>· Set clock time ( Year/Month/Date, Hour/Minute/ Second )</li> <li>· Set sampling time</li> <li>· Auto power OFF management</li> <li>· Set beep Sound ON/OFF</li> <li>· Decimal point of SD card setting</li> <li>· SD memory card Format</li> <li>· Set Metric/Imperial unit</li> </ul>		
Data Hold	Freeze the display reading. * Only available for the RMS function.		
Memory Recall	Maximum & Minimum value. * Only available for the RMS function.		
Data Output	RS 232/USB PC computer interface. * Connect the optional RS232 cable UPCB-02 will get the RS232 plug. * Connect the optional USB cable USB-01 will get the USB plug.		
Sampling Time of Display	Approx. 1 second.		
Operating Temperature and Humidity	0 to 50 °C. Less than 85% R.H.		
Power Supply	* Alkaline or heavy duty DC 1.5 V battery ( UM3, AA ) x 8 PCs, or equivalent.		
	* DC 9V adapter input. ( AC/DC power adapter is optional ).		
Power Current	Normal operation ( w/o SD card save data and LCD Backlight is OFF ) : Approx. DC 12 mA.		
	When SD card save the data and LCD Backlight is OFF ) :Approx. DC 35 mA.		
Weight	Meter : 515 g/ 1.13 LB.		
	Probe with cable and magnetic base : 99 g/0,22 LB		
Dimension	Meter : 203 x 76 x 38 mm		
	Vibration sensor probe: Round 16 mm Dia. x 37 mm. Cable length : 1.2 meter.		

Accessories Included	<ul style="list-style-type: none"> <li>* Instruction manual..... 1 PC</li> <li>* Vibration sensor set, VB-83 with cable..... 1 PC</li> <li>* Magnetic base.....1 PC</li> <li>* Carrying case ( CA-08 )..... 1 PC</li> </ul>
Optional Accessories	<ul style="list-style-type: none"> <li>* Vibration sensor set, VB-83 with cable.</li> <li>* SD Card ( 4 G ).</li> <li>* AC to DC 9V adapter.</li> <li>* USB cable, USB-01.</li> <li>* RS232 cable, UPCB-02.</li> <li>* Data Acquisition software,SW-U801-WIN.</li> </ul>

**ELECTRICAL SPECIFICATIONS (23± 5 °C)**

**Acceleration ( RMS, Peak, Max Hold )**

Unit	m/s <sup>2</sup>
Range	0.5 to 199.9 m/s <sup>2</sup>
Resolution	0.1 m/s <sup>2</sup>
Accuracy	± ( 5 % + 5 d ) reading @ 160 Hz, 80 Hz, 23 ± 5 °C
Calibration Point	50 m/S <sup>2</sup> ( 160 Hz )

Unit	g @ 1 g = 9.8 m/s <sup>2</sup>
Range	0.05 to 20.39 G
Resolution	0.01 G
Accuracy	± ( 5 % + 5 d ) reading @ 160 Hz, 80 Hz, 23 ± 5 °C
Calibration Point	50 m/S <sup>2</sup> ( 160 Hz )

Unit	ft/s <sup>2</sup>
Range	2 to 656 ft/s <sup>2</sup>
Resolution	1 ft/s <sup>2</sup>
Accuracy	± ( 5 % + 5 d ) reading @ 160 Hz, 80 Hz, 23 ± 5 °C
Calibration Point	50 m/S <sup>2</sup> ( 160 Hz )

Remark :

RMS : To measure the true RMS value.

Peak : To measure and update the peak value.

Max. Hold : To measure and update the max. peak value.

**Velocity ( RMS, Peak, Max Hold )**

Unit	mm/s
Range	0.5 to 199.9 mm/s
Resolution	0. 1 mm/s
Accuracy	± ( 5 % + 5 d ) reading @ 160 Hz, 80 Hz, 23 ± 5 °C
Calibration Point	50 mm/s ( 160 Hz )

Unit	cm/s
Range	0.05 to 19.99 cm/s
Resolution	0. 01 cm/s
Accuracy	± ( 5 % + 5 d ) reading @ 160 Hz, 80 Hz, 23 ± 5 °C
Calibration Point	50 mm/s ( 160 Hz )

Unit	inch/s
------	--------

Range	0.02 to 7.87 inch/s
Resolution	0.01 inch/s
Accuracy	$\pm ( 5 \% + 5 d )$ reading @ 160 Hz, 80 Hz, $23 \pm 5$ °C
Calibration Point	50 mm/s ( 160 Hz )
Remark : RMS : To measure the true RMS value. Peak : To measure and update the peak value. Max. Hold : To measure and update the max. peak value.	

***Displacement ( p-p, Max Hold p-p )***

Range	1.999 mm
Resolution	0.001 mm
Accuracy	$\pm ( 5 \% + 5 d )$ reading @ 160 Hz, 80 Hz, $23 \pm 5$ °C
Calibration Point	0.141 mm ( 160 Hz )

Unit	inch
Range	0.078 inch
Resolution	0.001 inch
Accuracy	$\pm ( 5 \% + 5 d )$ reading @ 160 Hz, 80 Hz, $23 \pm 5$ °C
Calibration Point	0.141 mm ( 160 Hz )
Remark : p-p : To measure the Peak to Peak value. Max. Hold p-p : To measure and update the max. Peak to Peak value.	

\* Appearance and specifications listed in this brochure are subject to change without notice.

1609-BVB8217SD

**www.YaLAB.com.tw    02-2389-0101**