

# Excel DATA ACQUISITION SOFTWARE for LCR-9183, LCR-9184

ISO-9001, CE, IEC1010

www.yalab.com.tw 02-2389-0101 04-2227-0088


Microsoft Excel - 3W-E803\_9184

H25

Ver. :120422

Please press Meter's 'RS-32' button first.

Comm. Port	Config / Start	Stop	Manual Input	Clear					
1	2	3	4	5	6	7	8	9	10
Date	Time	Pri. Mode	Frequency	Pri. Value	Pri. Unit	Sec. Mode	Sec. Value	Sec. Unit	
2012/7/9	13:34:54	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:34:55	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:34:56	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:34:57	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None	
2012/7/9	13:34:58	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None	
2012/7/9	13:34:59	C Mode	1kHz	2.177	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:00	C Mode	1kHz	2.177	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:01	C Mode	1kHz	2.183	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:02	C Mode	1kHz	2.183	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:03	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:35:04	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:35:05	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:35:06	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None	
2012/7/9	13:35:07	C Mode	1kHz	2.176	nF	Dissipation fact	0.0000	None	
2012/7/9	13:35:08	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:09	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:10	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:11	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:12	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:13	C Mode	1kHz	2.189	nF	Dissipation fact	0.0010	None	
2012/7/9	13:35:14	C Mode	1kHz	2.176	nF	Dissipation fact	0.0000	None	



www.YaLAB.com.tw

The Art of Measurement

for LCR-9183, LCR-9184

# Excel DATA ACQUISITION SOFTWARE

Model : SW-E803

[www.YaLAB.com.tw](http://www.YaLAB.com.tw) 02-2389-0101 04-2227-0088

## FEATURES

- \*SW-E803 can log out the LCR-9183 or LCR-9184 measurement value on Microsoft Office Excel, which makes you much easier to share, analyze, and manage the measurement results.
- \*Microsoft Office Excel installed on PC is required.
- \*Compatible with Microsoft OS Windows 98, Windows XP, Windows Vista, and Windows 7.
- \*Manual data recording with real time ( year /date/min/sec ) information, it is the useful incoming quality control tool to check the components quality of Inductor, Capacitor, Resistor.
- \*Automatic data recording with real time data acquisition system. Time data can be recorded as year/date/min/sec.
- \*Data sampling time: 1 to 3,600 seconds.
- \*Data can be utilized in Excel with general Excel functions, ex. drawing the data chart.

Comm. Port	Config / Start	Stop	Manual Input	Clear	Please press Meter's "RS232" button first.				
Ver. : 120422	Date	Time	Pri. Mode	Frequency	Pri. Value	Pri. Unit	Sec. Mode	Sec. Value	Sec. Unit
1	2012/7/9	13:34:54	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
2	2012/7/9	13:34:55	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
3	2012/7/9	13:34:56	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
4	2012/7/9	13:34:57	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None
5	2012/7/9	13:34:58	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None
6	2012/7/9	13:34:59	C Mode	1kHz	2.177	nF	Dissipation fact	0.0010	None
7	2012/7/9	13:35:00	C Mode	1kHz	2.177	nF	Dissipation fact	0.0010	None
8	2012/7/9	13:35:01	C Mode	1kHz	2.183	nF	Dissipation fact	0.0010	None
9	2012/7/9	13:35:02	C Mode	1kHz	2.183	nF	Dissipation fact	0.0010	None
10	2012/7/9	13:35:03	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
11	2012/7/9	13:35:04	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
12	2012/7/9	13:35:05	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
13	2012/7/9	13:35:06	C Mode	1kHz	2.191	nF	Dissipation fact	0.0020	None
14	2012/7/9	13:35:07	C Mode	1kHz	2.176	nF	Dissipation fact	0.0000	None
15	2012/7/9	13:35:08	C Mode	1kHz	2.184	nF	Dissipation fact	0.0010	None
16	2012/7/9	13:35:09	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None
17	2012/7/9	13:35:10	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None
18	2012/7/9	13:35:11	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None
19	2012/7/9	13:35:12	C Mode	1kHz	2.186	nF	Dissipation fact	0.0010	None
20	2012/7/9	13:35:13	C Mode	1kHz	2.189	nF	Dissipation fact	0.0010	None
21	2012/7/9	13:35:14	C Mode	1kHz	2.176	nF	Dissipation fact	0.0000	None
22	2012/7/9	13:35:14	C Mode	1kHz	2.176	nF	Dissipation fact	0.0000	None